

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00792	Calib. Date: 07/28/2020
Certification File No.:	00793	Calib. No.: 00031
Linearity File No.:	00794	Cert. Date: 07/28/2020
Solution File No.:	00811	Cert. No.: 00026
Sequential File No.:	00811	Lin. Date: 07/28/2020
		Lin. No.: 00026
		Soln. Date: 11/10/2020
		Soln. No.: 00202
		File Date: 11/10/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDZH-0064
Control Solution %:	0.100%		Expires: 02/19/2022
Solution Control Lot:	20090		Bottle No.: 0426

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:25S		
Control 1 EC	0.101%	16:25S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:25S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:26S		
Control 2 EC	0.100%	16:27S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:27S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27S		
Control 3 EC	0.100%	16:28S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:28S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

Lt. F. Pumphrey #3786

Badge No.: 3786

Date: 11/10/2020