

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXC-0091
Location:	CEDAR GROVE POLICE	
Calibration File No.:	00754	Calib. Date: 10/20/2019
Certification File No.:	00755	Calib. No.: 00029
Linearity File No.:	00756	Cert. Date: 10/20/2019
Solution File No.:	00779	Cert. No.: 00024
Sequential File No.:	00779	Lin. Date: 10/20/2019
		Lin. No.: 00024
		Soln. Date: 02/27/2020
		Soln. No.: 00190
		File Date: 02/27/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDZH-0063
Control Solution %:	0.100%		Expires: 04/09/2021
Solution Control Lot:	19110		Bottle No.: 0733

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:18S		
Control 1 EC	0.100%	12:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19S		
Control 2 EC	0.099%	12:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21S		
Control 3 EC	0.099%	12:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: _____

LT. F. Pumphrey #3786

Badge No.: 3786

Date: 02/27/2020