

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXC-0091  
Location: CEDAR GROVE POLICE  
Calibration File No.: 00880 Calib. Date: 09/13/2022 Calib. No.: 00036  
Certification File No.: 00881 Cert. Date: 09/13/2022 Cert. No.: 00031  
Linearity File No.: 00882 Lin. Date: 09/13/2022 Lin. No.: 00031  
Solution File No.: 00886 Soln. Date: 11/09/2022 Soln. No.: 00232  
Sequential File No.: 00886 File Date: 11/09/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZH-0064  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 0546

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:57S		
Control 1 EC	0.102%	18:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	18:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:59S		
Control 2 EC	0.101%	18:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:00S		
Control 3 EC	0.101%	19:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PUMPHREY

First Name: FRANCIS

MI: M

Signature: \_\_\_\_\_

*Capt. F. Pumphrey #3786*

Badge No.: 3786

Date: 11/09/2022